

# 256mV

ORP CALIBRATION SOLUTION

(25°C)  $\pm 15\text{mV}$

## ***Direction:***

\*Pour the solution in a clean container. Immerse the ORP probe tip into solution. After reading stabilizes, adjust the ORP calibration trimmer on the ORP meter until the display value shows 256mV or the reading is within the acceptable accuracy.

\*Never pour used solution back into this bottle.

**Expires in 10 days after opening**